

Below, please find abbreviated instructions, but remember to read the complete instructions that are packaged with your AEROECLIPSE® XL Reusable Breath Actuated Nebulizer (R BAN) device.

1 The AEROECLIPSE® XL R BAN can be used with a mask by removing the supplied mouthpiece and inserting the mask into the mouthpiece opening.

2 Untwist and remove the nebulizer top. Carefully place prescribed medication into nebulizer cup. Reattach the nebulizer top; gently hand tighten.

3 Attach one end of the supplied tubing to the fitting on the bottom of the nebulizer and the other end to your compressor. Be sure both ends of the tubing are secure.

4 Turn on your compressor while sitting in a comfortable, upright position. Place the mouthpiece in your mouth being sure not to cover the exhalation valve with your lips.

AEROECLIPSE® XL REUSABLE BREATH ACTUATED NEBULIZER (R BAN)



5 Inhale slowly and deeply. As you breathe in the green button on top of the nebulizer will move into the “down” position.

6 Exhale normally. The green button will return to the “up” position when you are not inhaling.

7 Continue to breathe normally on the nebulizer until it begins to make a “sputter” noise or as instructed by your physician or health care provider.

8 Turn off the compressor when finished with your treatment. Clean product between each treatment as directed in the owner’s manual.

This product is for single patient use only and has a 6 month life if properly used, cleaned and maintained.

A continuación se detalla un resumen de las instrucciones, pero recuerde leer las instrucciones completas que recibió con el nebulizador reutilizable activado por la respiración AeroEclipse® XL (R BAN).

1 El nebulizador AEROECLIPSE® XL R BAN puede utilizarse con una máscara. Para ello, quite la boquilla provista e introduzca la máscara en la abertura correspondiente.

2 Desenrosque y quite la parte superior del nebulizador. Coloque con cuidado el medicamento recetado en el recipiente del nebulizador. Vuelva a conectar la parte superior del nebulizador y apriete a mano con suavidad.

3 Enchufe un extremo del tubo provisto en el conector de la parte inferior del nebulizador y el otro extremo al compresor. Asegúrese de que ambos extremos de tubo estén firmes.

4 Siéntese en una posición cómoda, vertical y encienda el compresor. Colóquese la boquilla en la boca, pero tenga la precaución de no cubrir la válvula de exhalación con los labios.

NEBULIZADOR REUTILIZABLE ACTIVADO POR LA RESPIRACIÓN AEROECLIPSE® XL (R BAN)



5 Inhale lenta y profundamente. A medida que inhala, el botón de color verde en la parte superior del nebulizador se desplazará “hacia abajo”.

6 Exhale normalmente. El botón verde regresará a la posición “hacia arriba” cuando usted no esté inhalando.

7 Continúe respirando normalmente en el nebulizador hasta que comience a “chisporrotear” o según le indique el médico o profesional que le atiende.

8 Apague el compresor cuando finalice el tratamiento. Limpie el producto después de cada tratamiento, tal como se indica en el manual del propietario.

Este producto está diseñado para un solo paciente y tiene una vida útil de seis (6) meses si se utiliza, limpia y mantiene correctamente.

Las palabras o frases acompañadas por ™ o ® son marcas comerciales y marcas registradas de Monaghan Medical Corporation o de una filial de Monaghan Medical Corporation. © 2009-2020 Monaghan Medical Corporation. † marcas comerciales y marcas registradas de las respectivas empresas.